Calibration Certificate
PQ Systems

Date: 6/19/2009

Number:	C-01001	Manufacturer:	Fowler
Location:	Gage Room	Size:	1"
Item name:	Digi inside w absolute encoder	Serial number:	18732
Accuracy:	0		
Model:	FIC-23451		
Master gage(s):	MASTER-06001 MASTER-06002 MASTE		

Calibration interval: 6 Month Next due date: 12/11/2009

Certificate number:	Cert-0050
Gage condition:	Good
Temperature:	
Humidity:	
Calibrated on:	6/11/2009
Calibration done by:	SUPERVISOR
Calibration entered by:	SUPERVISOR
Adjustment made:	No
Overall calibration result:	Pass

Comment:

Name	Target	Acceptable +	Acceptable -	Initial result	After adjustment
Min-Range	0.100	0.102	0.098		
Mid-Range	0.500	0.502	0.498		
Max-Range	1.000	1.002	0.998		

GAGEpack® EZ