Calibration Certificate
PQ Systems

Date: 6/23/2009

Asset #:	C-01001	Manufacturer:	Fowler
Location:	Gage Room	Size:	1"
Item name:	Digi inside w absolute encoder	Serial number:	18732
Accuracy:	0		
Model:	FIC-23451		

Standards Used to Calibrate Equipment

Serial number	GageID	Last calibration	Due for calibration
10034	MASTER-06003	1/16/2009	1/15/2010

Calibration interval: 6 Month Next due date: 12/11/2009

Certificate number:	Cert-0050
Gage condition:	Good
Calibrated on:	6/11/2009
Calibration done by:	SUPERVISOR
Calibration entered by:	SUPERVISOR
Calibration lab manager:	Steve Richey
Adjustment made:	No
Overall calibration result:	Pass

Comment:

Name	Target	Units	Acceptable +	Acceptable -	Initial result
Min-Range	0.100	millimeters	0.102	0.098	
Mid-Range	0.500	millimeters	0.502	0.498	
Max-Range	1.000	millimeters	1.002	0.998	

GAGEpack® EZ