Calibration Certificate
PQ Systems Date:6/23/2009

Number:	C-01001	Manufacturer:	Fowler
Location:	Gage Room	Size:	1"
Item name:	Digi inside w absolute encoder	Serial number:	18732
Uncertainty:		Resolution:	0
Model:	FIC-23451	Procedure:	1. Carefully examine all gaging surfaces for ni
Master gage(s):	MASTER-06001 MASTER-06002 MASTE		

Compensating corrections: ( ) Yes ( ) No If Yes, describe them here:

Std/Master identification and traceability information

Calibration interval: 6 Month Next due date: 12/11/2009

Certificate number:	Cert-0050
Calibration meets requirements of:	ISO 10012
Gage condition:	Good
Temperature:	
Humidity:	
Calibrated on:	6/11/2009
Calibration done by:	SUPERVISOR
Calibration entered by:	SUPERVISOR
Adjustment made:	No
Calibration Time:	1.25
Overall calibration result:	Pass

Comment:

Name	Target	Units	Acceptable +	Acceptable -	Initial result
Min-Range	0.100	millimeters	0.102	0.098	
Mid-Range	0.500	millimeters	0.502	0.498	
Max-Range	1.000	millimeters	1.002	0.998	

GAGEpack® EZ