

Calibration results - Gages

PQ Systems

Printed: 6/23/2009

All Dates

Active Gages

Gage number: C-01001

Calib. date	Step name	Target value	As found	After adj.	Deviation
8/1/2001	Min-Range	0.1000	0.1000	0.1000	0.0000
	Mid-Range	0.5000	0.5000	0.5000	0.0000
	Max-Range	1.0000	1.0000	1.0000	0.0000
11/1/2001	Min-Range	0.10000	0.10050	0.10000	0.00000
	Mid-Range	0.50000	0.50030	0.50010	0.00010
	Max-Range	1.00000	1.00350	1.00090	0.00090
2/1/2002	Min-Range	0.10000	0.10300	0.10000	0.00000
	Mid-Range	0.50000	0.50010	0.50000	0.00000
	Max-Range	1.00000	1.00190	1.00000	0.00000
1/24/2003	Min-Range	0.1000	0.1000		
	Mid-Range	0.5000	0.5000		
	Max-Range	1.0000	1.0000		
5/28/2003	Min-Range	0.1000	0.0990		
	Mid-Range	0.5000	0.4990		
	Max-Range	1.0000	0.9980		
4/27/2004	Min-Range	0.100	0.100		
	Mid-Range	0.500	0.500		
	Max-Range	1.000	0.999		
4/29/2004	Min-Range	0.1000	0.1000		
	Mid-Range	0.5000	0.5000		
	Max-Range	1.0000	1.0000		
10/29/2004	Min-Range	0.1	0.1		
	Mid-Range	0.5	0.5		
	Max-Range	1.0	1.0		
4/29/2005	Min-Range	0.100	0.101		
	Mid-Range	0.500	0.501		
	Max-Range	1.000	1.001		
10/28/2005	Min-Range	0.1000	0.0999		
	Mid-Range	0.50000	0.49999		
	Max-Range	1.0000	0.9999		
4/28/2006	Min-Range	0.100	0.099		
	Mid-Range	0.500	0.501		
	Max-Range	1.000	1.000		
10/27/2006	Min-Range	0.100	0.100		
	Mid-Range	0.500	0.500		

